

TSKgel® G2000SW_{XL}/ BioAssist G2SW_{XL}/ QC-PAK GFC 200 Products

Part Numbers:	08540, TSKgel G2000SW _{XL} , 7.8 mm ID x 30 cm (stainless)	08543, Guard column for TSKgel G2000SW _{XL} , QC-PAK GFC 200 columns (stainless)
	20027, TSKgel BioAssist G2SW _{XL} , 7.8 mm ID x 30 cm (PEEK)	18008, Guard column for TSKgel BioAssist G2SW _{XL} column (PEEK)
	16214, QC-PAK GFC 200GL, 8 mm ID x 15 cm (glass)	
	16215, QC-PAK GFC 200, 7.8 mm ID x 15 cm (stainless)	08544, TSKgel SW _{XL} top-off gel, 1g

This sheet contains the recommended operating conditions and the specifications for TSKgel G2000SW_{XL}, BioAssist G2SW_{XL} and QC-PAK GFC 200 columns. Installation instructions and column care information are described in a separate Instruction Manual.

A. OPERATING CONDITIONS	
1. Shipping Solvent:	0.05% NaN ₃ and 0.1 mol/L Na ₂ SO ₄ in 0.1 mol/L phosphate buffer, pH 6.7
2. Max. Flow Rate:	1.2 mL/min
	When a buffer with high viscosity is used, the maximum flow rate may have to be reduced so as not to exceed the maximum pressure drop. When changing solvents, use a flow rate equal to 25% of the maximum flow rate.
3. Standard Flow Rate:	0.5 - 1.0 mL/min
4. Max. Pressure:	7.2 MPa (TSK-GEL G2000SW _{XL} , BioAssist G2SW _{XL}) 4.1 MPa (QC-PAK GFC 200)
5. pH Range:	2.5 - 7.5
6. Salt Conc.:	≤ 0.5 mol/L
7. Organic Conc.:	0 - 100% for aqueous soluble organic solvents. Make gradual solvent changes using a shallow gradient at low flow rate.
8. Temperature:	10 – 30 °C. Reduce flow rate when operating below 10 °C.
9. Cleaning Solvents:	1. Turn the column in reverse flow direction and run at half the maximum flow rate. 2. Clean with 5 column volumes (CV) of 1 mol/L sodium chloride, pH 7.0 3. Clean with 5CV of ultra-pure water. 4. Clean with 5CV of 20% acetonitrile. 5. Clean with 5CV of ultra-pure water. 6. Turn column in normal flow direction and equilibrate in mobile phase for at least 45 minutes
10. Storage:	Store the column in mobile phase containing 0.05% NaN ₃ or 20% ethanol when it will not be used the next day. For overnight storage flush the column with mobile phase at low flow rate. Prevent air from entering the column!
11. Column Protection:	The use of guard columns is recommended to prolong the life of the analytical column. Guard column life depends greatly on sample cleanliness. As a general rule, guard columns should be replaced after every 30-40 sample injections, when the peaks become excessively wide, or when the peaks show splitting.
12. TSKgel SW _{XL} top-off gel:	Occasionally, due to accident, sample, mobile phase or operational variables, a depression can develop at the column or guard column inlet. Use TSKgel SW _{XL} top-off gel for filling in such voids.
B. SPECIFICATIONS	
The performance of TSKgel G2000SW _{XL} , BioAssist G2SW _{XL} and QC-PAK GFC 200 columns are tested under the conditions described in the Data Sheet. All columns have passed the following quality control specifications:	
1. Theoretical Plates (N):	≥ 20,000 000 (TSKgek G2000SW _{XL} , BioAssist G2SW _{XL}) ≥ 10,000 (QC-PAK GFC 200)
2. Asymmetry Factor (AF):	0.7 - 1.6