## **RESIN INFORMATION SHEET**

Product Name	TOYOPEARL® GigaCap DEAE-650M
--------------	------------------------------

(Weak anion exchange chromatography resin)

Part Numbers 22865 TOYOPEARL GigaCap DEAE-650M, 100 mL

 22866
 TOYOPEARL GigaCap DEAE-650M, 250 mL

 22867
 TOYOPEARL GigaCap DEAE-650M, 1 L

 22868
 TOYOPEARL GigaCap DEAE-650M, 5 L

 22869
 TOYOPEARL GigaCap DEAE-650M, 50 L

#### **Product Description**

**TOYOPEARL** chromatographic resins are based on a rigid methacrylic polymer, resulting in high mechanical and chemical stability. Resins are available as non-functionalized "HW" series resins for size exclusion separations, and derivatized with surface chemistries for alternative modes of chromatography such as ion exchange, hydrophobic interaction or affinity separations.

**TOYOPEARL GigaCap DEAE-650M** chromatographic resins are designed for anion exchange chromatography. This chromatographic mode separates molecules on the basis of ionic interactions between the sample and the resin. The separation is usually accomplished in buffered aqueous solution with a gradient of increasing ionic strength. Alternatively, pH adjustment may be used for control of elution.

## Operating Conditions

Packing pressure	Typically 0.3 MPa
Shipping solvent	20 % (v/v) ethanol
Shipping formulation	72 % (v/v) slurry in shipping solvent (*)
Pressure limiting factor	Depend on column hardware (typically 0.7 MPa)
Operating linear flowrate	Typically 10 - 600 cm/hour
Long-term storage conditions	20 % (v/v) ethanol
Cleaning-in-place/Sanitization	0.1 - 0.5 mol/L NaOH or 0.1 mol/L HCl

### **Specifications**

50 - 100 μm
0.15 - 0.25 eq/L
Min. 156 g/L
Max. 100 CFU/mL
Max. 10.0 EU/mL
Max. 0.2 % (for dry gel)
Max. 6

# Additional Information

	Appearance	White resin slurry which settles upon standing
•	Mean pore diameter (base resin)	100 nm (*)
-		(A) T1

<sup>(\*)</sup> The value is for reference only, not guaranteed.

Lot-specific data are included in the Certificate of Analysis (COA) shipped with the product. For detailed test procedures please refer to the appropriate Regulatory Support File.